

## INTRODUCTION

### INTENDED USE

The InPouch® TV is a self-contained media system for the recovery and detection of *T. vaginalis* from female vaginal samples or male urethra/urine samples. The proprietary medium is selective for the transport and growth of *T. vaginalis*, while inhibiting the growth of other microorganisms, which can interfere with a reliable diagnosis.

### EXPLANATION

Human trichomoniasis is a sexually transmitted infection (STI) caused by the flagellated protozoan, *Trichomonas vaginalis*. It is recognized as one of the most prevalent sexually transmitted infections world-wide, in both males and females.<sup>1,2</sup> The CDC estimates five million new cases occur in the U.S. annually.

### PRINCIPLES OF THE PRODUCT

The pouch is designed for user-friendly and convenient early microscopic detection by culture confirmation of *T. vaginalis*.<sup>3,4</sup> The pouch consists of a high-barrier, oxygen-resistant, plastic with two V-shaped-chambers connected by a narrow passage that, together, provide a variety of benefits. The pouch allows users to easily inoculate a specimen, immediately observe (wet mount) the specimen, store and/or transport (optional) before transfer to the lab for incubation and recording.

### REAGENTS

The InPouch® medium contains the following: peptones, maltose and other sugars, amino acids, salts and antimicrobial agents in a phosphate buffered saline base. An unopened pouch should contain a clear, amber liquid.

### SPECIFICITY/SENSITIVITY

For cultivation of *T. vaginalis* only.

Clinical Specificity: 100%<sup>5</sup>; Clinical Sensitivity: 81-94%.<sup>5,6</sup>

### HANDLING

**Handle all inoculated tests in accordance to CDC-NIH recommendation for BSL-2 organisms.**

### STORAGE & SHELF LIFE

- Store uninoculated pouches at room temperature (18°C-25°C) horizontally and away from direct sunlight.
- **Never refrigerate or freeze the product.**
- InPouch® products expire 12 months from the date of manufacture.

## USING THE TEST

Gather  
Materials

#### Materials needed:

- InPouch® TV test (s) - (Supplied)
- Product Insert - (Supplied)
- Microscope Viewing Clip (**Catalog #14-2205, #14-2210**)
- Laboratory incubator
- Microscope with 10x/20x/40x objectives
- Disposable gloves
- Disposable pipette (urine sediment)
- Disposable cotton or rayon swabs (inoculation of the pouch from UTM, Amies or ESwab™ within 24 hours of sampling at room temperature are acceptable<sup>10,11</sup>)

Collect  
Sample

**Vaginal:** use a sterile cotton or rayon swab to collect a specimen from the posterior fornix and inoculate immediately.

**Urine and CSF:** Centrifuge a fresh specimen (less than 1 hour old) in a sterile disposable tube at 500×G for 5 minutes. Decant the supernatant. Inoculate the pouch.

**Seminal Fluid:** the specimen should be ≤ 60 minutes old. Use a disposable pipette to collect a drop of SF and inoculate the pouch.

Inoculate the  
pouch

To avoid fluid leakage, squeeze the fluid from the top of the pouch downward into the bottom chamber. Tear off the plastic top above the white closure. To admit the swab, open the pouch by pulling the closure tape middle tabs apart. Carefully squeeze a small amount of liquid back into the top chamber and knead the swab between the pouch walls. Remove swab and discard.

Immediate  
Observation

Before expressing sample to lower chamber and prior to incubation, roll the top edge down *at least twice* and secure the end tabs to seal the pouch. Isolate the sample in the Viewing Clip and observe it in the top chamber microscopically. Observation of live motile trichomonads is a presumptive positive result. Submit the sample to a lab for confirmation.

Incubation  
& Transport

**Express the sample and liquid to the bottom chamber.** Tightly roll the open end of the upper chamber over 2 to 3 times and seal by bending the tabs around the side of the pouch.

**NOTE: Fill in the patient information and place patient label over the blue Biomed label, not on the viewing chamber.**

*Incubate the pouch vertically at 37° C for up to 7 days.<sup>8</sup> The pouch is designed for safe transport.<sup>9</sup> Inoculated tests should be*

*transported within 48 hours after inoculation and maintained at 18° - 37°C.<sup>7</sup>*


## READING THE RESULTS

### MICROSCOPIC EVALUATION

To search for the presence of trichomonads, place a Viewing Clip horizontally over the lower chamber of the pouch and close (the Viewing Clip is optional). Place the pouch on the microscope stage under low power (100x) to look for trichomonads. Use a higher power (200x - 400x) if necessary for confirmation.



**Observation of 1 or more live, *T. vaginalis* cells is all that is required for a presumptive positive result. Continue incubation and microscopic observation daily for 5 working days before a negative result is reported.<sup>8</sup>**



### READING TIPS

- \* Trichomonads gravitate to the bottom and side edges of the pouch chamber.
- \* Verify that your field of focus is in the liquid and not the textured plastic film layer of the pouch.

### ADDITIONAL PRODUCT NOTES

1. NEVER refrigerate or freeze the pouch.
2. Complete each label with the patient information.

## DISPOSAL & SAFETY

Because inoculated pouches potentially contain live, infectious organisms, the pouch must be autoclaved at 121° C for 20 minutes or by other means appropriate for sterilization and disposal of BSL-2 organisms.

The InPouch® growth medium suppresses but may not entirely eliminate yeast and bacterial growth. Any build-up of gas from bacterial growth can be vented by opening the pouches inside a BSL-2 rated biological safety cabinet.

**WARNING:** *This product contains chemicals known to the State of California to cause cancer, birth defects and other reproductive harm.*

## QUALITY CONTROL

The InPouch® TV product is for presumptive human *T. vaginalis* identification only. The InPouch® TV product is manufactured in accordance with controlled procedures at Biomed Diagnostics. Each lot undergoes an initial QC performance testing prior to release for customer use. Additional performance testing is repeated throughout the marked shelf-life of each lot to ensure absolute reliability.

The following is recommended for customers who choose to complete independent QC testing of the InPouch® TV product:

1. Obtain a sample of viable *T. vaginalis* organisms in the range of  $2.0 \times 10^6$  live cells/mL.
2. Inoculate three (3) InPouch® tests with 1-12 drops (20 – 40 µL) of the live culture using a sterile glass Pasteur pipette per the “Inoculate the InPouch®” step as described on this insert.
3. Incubate the inoculated InPouch® tests for 24 hours at 37°C. After incubation, re-suspend the sample by kneading the pouch. Examine each pouch microscopically (10x objective) and confirm that you have viable *T. vaginalis* organisms in the range of  $2.0 \times 10^3$  to  $2.0 \times 10^6$  live cells/mL. Incubate the pouches for an additional 24 hours if necessary to confirm the doubling time.

## NOTES ON QUALITY

1. Menses does not interfere with the test.
2. There are no patient age limitations on specimens collected.

## *Trichomonas vaginalis* LIVE CULTURE

Live cultures of *T. vaginalis* (clinical isolate) for research, training and QC purposes are available (N. American customers only). This live culture (positive control) can be purchased from Biomed Diagnostics (Catalog #11-2015) to obtain an active culture of *T. vaginalis*.

## TECHNICAL NOTES

### An Evaluation of *Trichomonas vaginalis* Culture Viability After 48 Hours at Room Temperature

*Trichomonas vaginalis* SJCR66 was incubated for 48 hours at 37°C in an InPouch® TV test pouch. A Neubauer hemocytometer was used to determine a final dilution of the culture to  $7.75 \times 10^4$  cells/mL.

## TECHNICAL NOTES, CONTINUED

Four pouches were then inoculated from this dilution of trichomonads. Pouch #1 was inoculated with 30 µL, Pouch #2 with 60 µL, Pouch #3 with 90 µL and Pouch #4 with 120 µL.

The pouch densities were:

Pouch #1  $2.3 \times 10^3$  cells/mL Pouch #3  $6.9 \times 10^3$  cells/mL  
Pouch #2  $4.6 \times 10^3$  cells/mL Pouch #4  $9.2 \times 10^3$  cells/mL

All four pouches were allowed to remain at room temperature for 48 hours before being placed into an incubator at 37°C for 24 hours. Each pouch was examined microscopically for viability after 24 hours. All were positive for motile trichomonads with Pouch #1 presenting the fewest and Pouch #4 the greatest numbers of organisms. This demonstrates that a specimen inoculated into the pouch containing viable trichomonads at  $2.3 \times 10^3$  cells/mL would remain viable for a minimum of 48 hours if stored at room temperature.

Kenneth A. Borchardt, Ph.D., Chief Research Scientist Doc. No. 100-045B

## REFERENCES

1. Krieger JN, et al., JAMA 1988; 259:1223-1227
2. Krieger JN, Trichomoniasis in Men: Old Issues and New Data. *Sexually Transmitted Diseases*. 1995; 22:83-96.
3. Draper D, et al., Detection of *T. vaginalis* in pregnant women with the InPouch TV culture system, *J Clinical Microbiology*. 1993; 31:1016-1018.
4. Borchardt KA, et al., A Clinical Evaluation of Trichomoniasis in San Jose, Costa Rica using the InPouch TV test. *Genitourin Med*. 1992; 68:328-330.
5. Beverly AL, et al., Viability of *T. vaginalis* in transport medium. *J Clinical Microbiology*. 1999; 37: 37-49
6. Ohlemeyer CL, et al., Diagnosis of *Trichomonas vaginalis* in adolescent females: InPouch TV culture versus wet-mount microscopy. *Journal of Adolescent Health*, 1998; 22(3): 205-208
7. Huppert JS, et al., Rapid antigen testing compares favorably with transcription-mediated amplification assay for the detection of *T. vaginalis* in young women. *Clinical Infectious Diseases*. 2007; 45(2):194-198.
8. Rivers CA, et al., Diagnostic Rates Differ on Basis of the Number of Read Days with the Use of the InPouch Culture System for *Trichomonas vaginalis* Screening. *J Clinical Microbiology*. 2013; 51(11):3875-3876
9. Barenfanger J, et al., Timing and inoculation of the pouch makes no difference in increased detection of *T. vaginalis* by the InPouch Method. *J Clinical Microbiology*. 2002; 40:1387-1389.
10. Rivers CA, et al., Viability of *Trichomonas vaginalis* in Copan Universal Transport Medium and eSwab Transport Medium. *J Clinical Microbiology*. 2008; 46(9): 3134-3135. Beverly AL, et al., Viability of *Trichomonas vaginalis* in Transport Medium. *J Clinical Microbiology*. 1999; 37(11):3749-3750.

# InPouch® TV

## *Trichomonas vaginalis* Test

Catalog No. 10-2010 10 Pack  
Catalog No. 10-2003 100 Pack

A SELECTIVE CULTURE SYSTEM FOR  
THE DIAGNOSIS OF HUMAN

*Trichomonas vaginalis*

For *In Vitro* Diagnostic Use Only



# BIOMED

Manufactured by:

Biomed Diagnostics, Inc.

PO Box 2366 • White City, OR 97503  
tel. (800)-964-6466 • fax. (541) 830-3001  
info@biomeddiagnostics.com  
[www.biomeddiagnostics.com](http://www.biomeddiagnostics.com)