



Certificate of Analysis
InPouch™ TF *Tritrichomonas foetus* Test

Product Number 11-1001, 11-1003

Lot Number: 5CA039

Expiration Date: 26 MAR 2016

Test	Specification	Result
Appearance	Clear, Straw to Amber	Amber
pH Test	6.6 to 6.8	6.69
Sterility	Aseptic Medium Production and Dispensing	Sterile
Doubling Time	3 – 8 Hours	5.47

QC Acceptance Date: 06 APR 2015

Signature: 