



**Certificate of Analysis**

**TF-Transit Tube**

Product Number 60-1010, 60-1050

Lot Number: **5DA039A**

Expiration Date: **26 MAR 2016**

Test	Specification	Result
Appearance	Clear, Straw to Amber	<b>Amber</b>
pH Test	6.7 ± 0.1	<b>6.69</b>
Packaging	Transit Tubes conform to <i>ANSI/AAMI ISO11135</i> standards.	<b>Sterile</b>
Filling	Aseptic Medium Production and Dispensing	<b>Sterile</b>

**QC Acceptance Date: 14 APR 2015**

**Signature:**

A handwritten signature in blue ink, appearing to be "C. R. Brown", is written over a horizontal line.