



Certificate of Analysis
InTray™ DM Dermatophyte Test

Product Number: 10-4001, 10-4007, 10-4007-E, CA0827005FB, CA0827020FK

Lot Number: **5DE0494**

DOM: **28 APR 2015**

Expiration Date: **28 JUL 2017**

Test	Specification	Results
Appearance	Clear, yellow	Clear yellow
Sterility	No growth on negative control	No growth
pH	5.6 ± 0.1	5.62
Microbiology (ATCC®):	Colony Appearance	
<i>T. mentagrophytes</i> 9533	Growth w/ color change ≤ 7 days	2 days
<i>T. rubrum</i> 28188	Growth w/ color change ≤ 7 days	5 days
<i>M. gypseum</i> 14683	Growth w/ color change ≤ 7 days	2 days
<i>A. niger</i> 16404	No growth / significant inhibited	No growth
<i>E. coli</i> 25922	No growth	No growth

QC Acceptance Date: 11 MAY 2015

Signature: 