



Certificate of Analysis
InPouch™ TV *Trichomonas vaginalis* Test

Product Number 10-2010 10-2003

Lot Number: **5EA055A**

Expiration Date: **30 APR 2016**

Test	Specification	Result
Appearance	Clear, Straw to Amber	Clear, amber
pH Test	6.05 to 6.15	6.08
Sterility	Aseptic Medium Production and Dispensing	Sterile
Doubling Time	3 – 8 Hours	4.80

QC Acceptance Date: 11 MAY 2015

Signature: 