



**Certificate of Analysis**

**InTray™ GC Agar**

Product Number: 10-8001, 10-8007

Lot Number: **5EA065**

Expiration Date: **14 MAY 2016**

Test	Specification	Results
Appearance	Chocolate brown, opaque	<b>Chocolate, opaque</b>
Sterility	No growth on negative control	<b>No growth</b>
pH	7.2 ± 0.2	<b>7.2</b>
<b>Microbiology (ATCC®):</b>	<b>Colony Appearance</b>	
<i>N. gonorrhoeae</i> 43069	Colony growth	<b>Growth</b>
<i>N. meningitidis</i> 13090	Colony growth	<b>Growth</b>
<i>N. sicca</i> 9913	No Growth / Significant inhibition	<b>No growth</b>
<i>E. coli</i> 25922	No Growth / Significant inhibition	<b>No Growth</b>
<i>P. mirabilis</i> 43071	No Growth / Significant inhibition	<b>Significant inhibition</b>
<i>S. epidermidis</i> 12228	No Growth / Significant inhibition	<b>No growth</b>
<i>C. albicans</i> 60193	No Growth / Significant inhibition	<b>No growth</b>

**QC Acceptance Date: 19 MAY 2015**

Signature: