



Certificate of Analysis

TF-Transit Tube

Product Number 60-1010, 60-1050

Lot Number: **5HA109B**

Expiration Date: **23 July 2016**

Test	Specification	Result
Appearance	Clear, Straw to Amber	Clear, amber
pH Test	6.7 ± 0.1	6.74
Packaging	Transit Tubes conform to <i>ANSI/AAMI ISO11135</i> standards.	Sterile
Filling	Aseptic Medium Production and Dispensing	Sterile

QC Acceptance Date: 05 August 2015

Signature: 