



# Introduction

## Intended Use

InTray® SAB-FungID™ w/ CC contains Sabouraud's dextrose agar with chloramphenicol and cycloheximide, a selective medium used to aid in the detection of dermatophyte fungi from clinical specimens with mixed microbiota.

## Description and Principle

InTray SAB-FungID w/ CC is a dynamic system with built-in components and features that are designed for user compatibility and ease of dermatophyte fungi detection (i.e., 10x objective microscopy direct from the tray). The medium allows for the growth and observation of distinct colony morphology and color (i.e., pigments) of dermatophyte fungi, while inhibiting most gram-positive bacteria, gram-negative bacteria, yeast and saprophytic fungi. Dermatophytes are fungi in the genera *Microsporum*, *Trichophyton* and *Epidermophyton*. They are capable of metabolizing keratin found in skin, hair and nails of living hosts. The fungi characteristically may invade the cutaneous tissue of the living host but rarely penetrate the subcutaneous tissue. Tinea and ringworm are two terms commonly used to describe dermatophytes.

## Reagents and Appearance

SAB-FungID w/ CC appears transparent with a light amber hue and contains peptic/casein digest, dextrose, chloramphenicol (0.050 g/L) and cycloheximide (0.40 g/L) with a final pH of 5.6 ± 0.2 at 25°C.

## Precautions, Safety and Disposal

For In Vitro Diagnostic Use. For professional use only.

Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing and gloves.

Once the tray has been inoculated and resealed, re-open only in a biological safety cabinet. Because of the potential for containing infectious materials, the tray must be destroyed by autoclaving at 121°C for 20 minutes.

## Storage

On receipt, store trays at 2-25°C in the dark. Avoid freezing or prolonged storage at temperatures greater than 40°C. Do not use trays if the medium shows signs of deterioration or contamination.

## Shelf Life

InTray SAB-FungID w/ CC expires 12 months from date of manufacture.

# Procedure

## Key Notes Regarding Specimen Collection

Specimen collection poses a major uncertainty in using this device.

**NAILS** - Often collecting viable material from infected nails is difficult because the living organisms are well under the nail itself. For best results, cut nails into small pieces.

**HAIR** - Samples should be grasped at the uninfected end and several (3-6) small pieces, about 2 cm long, should be cut from the infected portion for inoculation onto the medium.

**SKIN** - Scrapings should be taken with an inoculation tool that has been moistened with the medium or a sharp blade from the outer ridge of an active lesion. Vesicular fluid is unacceptable for dermatophyte culture. If the infected area is vesiculated, skin scrapings should be taken from the surface.

## Prepare Sample:

Use aseptic technique during specimen collection and handling. Remove any soap residue from the sampling area. Clean the area with 70% alcohol and permit to air dry.

## Collect Sample:

InTray SAB-FungID w/ CC is designed for culturing hair, skin and nail samples (i.e., cuttings/scrapings). All specimens should be handled according to CDC infectious materials isolation guidelines; [cdc.gov/infectioncontrol/guidelines/isolation](https://www.cdc.gov/infectioncontrol/guidelines/isolation)

## Materials Provided

- InTray SAB-FungID w/ CC test(s)

## Materials Required but Not Provided

- Sterile inoculation tool (e.g., cotton swab/forceps/scalpel blade)
- Laboratory incubator capable of incubation at 25-30°C

## 1 Prepare InTray



Allow tray(s) to warm to 18-25°C before use. Immediately label the tray with patient/sample information and date. Pull back the lower right corner adjacent to the clear window of the tray label until the protective seal is completely visible.

Remove the seal by pulling the tab. Discard the seal.

**DO NOT REMOVE OR ALTER THE WHITE FILTER STRIP OVER THE VENT HOLE!**

## 2 Inoculate Sample



Inoculate the specimen on the center surface of the medium. A sterile inoculating loop that has been moistened by touching the surface of the medium may be used for inoculation of solids or scrapings.

**Re-seal all around the tray to ensure a complete seal by pressing the edges of the label against the plastic tray. DO NOT COVER THE VIEWING WINDOW. Complete re-seal prevents dehydration!**

## Incubation

Incubate inoculated trays in a dark environment for up to 21 days at 25-30°C. Observe the trays daily through the clear viewing window.

## Quality Control

This product has been tested and meets the CLSI (formerly NCCLS) Approved Standard for commercially prepared media (M22-A3). At the time of manufacture, quality control testing is performed on each lot of the InTray SAB-FungID w/ CC. The ability of the media to support growth and demonstrate expected biochemical reactions and morphology is verified by lot.

## Strains for QC Testing SAB-FungID w/ CC

Organism	ATCC®	Expected Result
<i>T. mentagrophytes</i>	9533	Growth
<i>T. rubrum</i>	28188	Growth
<i>C. albicans</i>	60193	Partial Inhibition
<i>A. brasiliensis</i>	16404	Significant Inhibition
<i>S. aureus</i>	25923	Significant Inhibition
<i>E. coli</i>	25922	Significant Inhibition