



Certificate of Analysis

Timoney's CEM with 10% Horse Blood Agar

Product Number: 12-143-001, 12-143-002, 12-144-001, 12-144-002

Lot Number: **5MA182**

Expiration Date: **12JUN2026**

Test	Specification	Results
Appearance	Chocolate/Red	Chocolate/Red
Microbial load	No growth	No growth
pH	7.0 ± 0.1	7.05

Microbiology (STRAINS):		Colony Appearance	Results
<i>T. equigenitalis</i>	35865	Growth/oxichrome positive	Growth/oxichrome positive
<i>E. coli</i>	25922	Complete to partial inhibition, oxichrome negative	Partial inhibition, oxichrome negative
<i>S. xylosus</i>	29971	Complete to significant inhibition, oxichrome negative	Complete inhibition, oxichrome negative
<i>C. albicans</i>	90028	Complete to significant inhibition, oxichrome variable	Complete inhibition, oxichrome negative

QC Acceptance Date: 17DEC2025

Signature: *Megan Slama*