



Certificate of Analysis
InPouch™ TF *Tritrichomonas foetus* Test

Product Number 12-011-001, 12-011-002

Lot Number: 6RA043B

Expiration Date: 02APR2027

Test	Specification	Result
Appearance	Clear, Straw to Amber	Clear, Straw
pH Test	6.6 to 6.8	6.79
Microbial load	No growth	No growth
Doubling Time	3 – 8 Hours	4.68 Hours

QC Acceptance Date: 26MAY2026

Signature: *Megan Slama*