



Certificate of Analysis
InTray™ Mueller Hinton Chocolate Agar

Product Number: 12-483-001, 12-483-002

Lot Number: 6RA049

Expiration Date: 22APR2027

Test	Specification	Results
Appearance	Opaque, chocolate brown	Opaque, chocolate brown
Microbial load	No growth	No growth
pH	7.3 ± 0.2	7.19
Microbiology (ATCC®):	Colony Appearance	
<i>H. influenzae</i> 10211	Good growth	Good growth
<i>N. meningitidis</i> 13090	Good growth	Good growth
<i>S. pneumoniae</i> 6305	Good growth	Good growth

QC Acceptance Date: 15MAY2026

Signature: *Megan Slama*