



Certificate of Analysis
InTray™ GC Agar

Product Number: 11-080-001, 11-080-002

Lot Number: 6SA063

Expiration Date: 14MAY2027

Test	Specification	Results
Appearance	Chocolate brown, opaque	Chocolate brown, opaque
Microbial load	No growth	No growth
pH	7.2 ± 0.2	7.10
Microbiology (STRAIN):	Colony Appearance	
<i>N. gonorrhoeae</i> AR-167	Colony growth	Colony growth
<i>N. meningitidis</i> 13090	Colony growth	Colony growth
<i>N. sicca</i> 9913	No Growth / Partial Inhibition	No Growth
<i>E. coli</i> 25922	No Growth / Partial Inhibition	No Growth
<i>P. mirabilis</i> 43071	No Growth / Partial Inhibition	No Growth
<i>S. epidermidis</i> 12228	No Growth / Partial Inhibition	No Growth
<i>C. albicans</i> 60193	No Growth / Partial inhibition	Partial Inhibition

QC Acceptance Date: 21MAY2026

Signature: *Megan Slama*