



**Certificate of Analysis**  
**InTray™ COLOREX™ Screen**

Product Number: 11-103-001, 11-103-002

Lot Number: 6SA066

Expiration Date: 20MAY2027

Test	Specification	Results
Appearance	Opalescent, clear	Opalescent, clear
Microbial load	No growth	No growth
pH	7.0 ± 0.2	6.96
<b>Microbiology (ATCC®):</b>	<b>Colony Appearance</b>	
<i>E. coli</i> 25922	Rose	Rose
<i>K. pneumoniae</i> 13883	Metallic blue	Metallic blue
<i>P. mirabilis</i> 43071	Orange/brown, swarming	Orange/brown, swarming
<i>S. aureus</i> 25923	Opaque cream	Opaque cream
<i>S. sapro</i> 15305	Pale pink	Pale pink
<i>E. faecium</i> 6569	Turquoise/ blue	Turquoise/blue

**QC Acceptance Date: 27MAY2026**

**Signature:** *Megan Slama*