



Certificate of Analysis
InPouch™ TV *Trichomonas vaginalis* Test

Product Number 11-031-001, 11-031-002

Lot Number: 6QA036C

Expiration Date: 13MAR2027

Test	Specification	Result
Appearance	Clear, Straw to Amber	Clear, Straw
pH Test	6.05 to 6.15	6.09
Microbial load	No growth	No growth
Doubling Time	3 – 8 Hours	3.53 Hours

QC Acceptance Date: 05JUN2026

Signature: *Megan Slama*