



Certificate of Analysis

InTray™ COLOREX™ Screen

Product Number: 11-103-001, 11-103-002

Lot Number: 6SA066

Expiration Date: 20MAY2027

Test	Specification	Results
Appearance	Opalescent, clear	Opalescent, clear
Microbial load	No growth	No growth
pH	7.0 ± 0.2	6.96
Microbiology (ATCC®):		Colony Appearance
<i>E. coli</i>	25922	Rose
<i>K. pneumoniae</i>	13883	Metallic blue
<i>P. mirabilis</i>	43071	Orange/brown, swarming
<i>S. aureus</i>	25923	Opaque cream
<i>S. sapro</i>	15305	Pale pink
<i>E. faecium</i>	6569	Turquoise/blue

QC Acceptance Date: 27MAY2026

Signature: *Megan Slama*